

Certificado No. (Certificate #):62478Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-23Marca (Brand):INSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
275.00 mm	275.00 mm	0 µm	± 9 µm	±	5.8E+00 µm
276.00 mm	276.00 mm	0 µm	± 9 µm	±	5.8E+00 µm
277.00 mm	277.00 mm	0 µm	± 9 µm	±	5.8E+00 µm
278.00 mm	278.00 mm	0 µm	± 9 µm	±	5.8E+00 µm
279.00 mm	279.00 mm	0 µm	± 9 µm	±	5.8E+00 µm
280.00 mm	280.00 mm	0 µm	± 9 µm	±	5.8E+00 µm