

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	14.9 °C	-0.221 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.015 °C	19.9 °C	-0.115 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.0 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.447 %HR	20.1 %HR	-0.347 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	50.1 %HR	-0.613 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.775 %HR	80.3 %HR	-0.475 %HR	± 3.0 %HR	± 1.9E+00 %HR