

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04995	0.05		± 4.3E-01 mm					
0 a 0.1	0.09993	0.07		± 4.3E-01 mm					
0 a 0.15	0.15002	-0.02		± 4.3E-01 mm					
0 a 0.5	0.49980	0.20		± 4.3E-01 mm					
0 a 1	1.00035	-0.35		± 4.3E-01 mm					
0 a 2	1.99990	0.10		± 4.8E-01 mm					
0 a 3	2.99982	0.18		± 4.8E-01 mm					
0 a 4	3.99988	0.12		± 4.8E-01 mm					
0 a 5	4.99990	0.10		± 4.8E-01 mm					
0 a 6	5.99986	0.14		± 3.6E-01 mm					
0 a 7	7.00012	-0.12		± 3.6E-01 mm					
0 a 8	7.99974	0.26		± 3.6E-01 mm					
0 a 9	8.99980	0.20		± 3.6E-01 mm					
0 a 10	9.99964	0.36		± 3.6E-01 mm					
0 a 11	10.99970	0.30		± 3.6E-01 mm					
0 a 12	11.99957	0.43		± 3.6E-01 mm					
0 a 13	12.99960	0.40		± 3.6E-01 mm					
0 a 14	13.99950	0.50		± 3.6E-01 mm					
0 a 15	14.99954	0.46		± 3.6E-01 mm					
0 a 16	15.99945	0.55		± 3.6E-01 mm					
0 a 17	16.99941	0.59		± 3.6E-01 mm					
0 a 18	17.99949	0.51		± 3.6E-01 mm					
0 a 19	18.99958	0.42		± 3.6E-01 mm					
0 a 20	19.99961	0.39		± 3.6E-01 mm					
0 a 21	20.99961	0.39		± 5.5E-01 mm					
0 a 22	21.99960	0.40		± 5.5E-01 mm					
0 a 23	22.99964	0.36		± 5.5E-01 mm					
0 a 24	23.99964	0.36		± 5.5E-01 mm					
0 a 25	24.99961	0.39		± 5.5E-01 mm					
0 a 26	25.99964	0.36		± 5.5E-01 mm					
0 a 27	26.99961	0.39		± 5.5E-01 mm					
0 a 28	27.99964	0.36		± 5.5E-01 mm					
0 a 29	28.99966	0.34		± 5.5E-01 mm					
0 a 30	29.99961	0.39		± 5.5E-01 mm					