

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04967	0.330		± 4.3E-01 mm					
0 a 0.1	0.09964	0.360		± 4.3E-01 mm					
0 a 0.15	0.14987	0.130		± 4.3E-01 mm					
0 a 1	0.99953	0.470		± 4.3E-01 mm					
0 a 2	1.99972	0.280		± 4.8E-01 mm					
0 a 3	2.99950	0.500		± 4.8E-01 mm					
0 a 4	3.99959	0.410		± 4.8E-01 mm					
0 a 5	4.99950	0.500		± 4.8E-01 mm					
0 a 6	5.99950	0.500		± 3.6E-01 mm					
0 a 7	6.99972	0.280		± 3.6E-01 mm					
0 a 8	7.99970	0.300		± 3.6E-01 mm					
0 a 9	8.99982	0.180		± 3.6E-01 mm					
0 a 10	9.99989	0.110		± 3.6E-01 mm					
0 a 11	10.99991	0.090		± 3.6E-01 mm					
0 a 12	12.00022	-0.220		± 3.6E-01 mm					
0 a 13	13.00006	-0.060		± 3.6E-01 mm					
0 a 14	14.00039	-0.390		± 3.6E-01 mm					
0 a 15	14.99955	0.450		± 3.6E-01 mm					
0 a 16	15.99945	0.550		± 3.6E-01 mm					
0 a 17	17.00007	-0.070		± 3.6E-01 mm					
0 a 18	18.00029	-0.290		± 3.6E-01 mm					
0 a 19	19.00019	-0.190		± 3.6E-01 mm					
0 a 20	19.99995	0.050		± 3.6E-01 mm					
0 a 21	20.99991	0.090		± 5.5E-01 mm					
0 a 22	21.99989	0.110		± 5.5E-01 mm					
0 a 23	23.00030	-0.300		± 5.5E-01 mm					
0 a 24	24.00049	-0.490		± 5.5E-01 mm					
0 a 25	25.00050	-0.500		± 5.5E-01 mm					
0 a 26	26.00039	-0.390		± 5.5E-01 mm					
0 a 27	27.00029	-0.290		± 5.5E-01 mm					
0 a 28	28.00004	-0.040		± 5.5E-01 mm					
0 a 29	29.00022	-0.220		± 5.5E-01 mm					
0 a 30	29.99977	0.230		± 5.5E-01 mm					