

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
2 mm	1.90 mm	-100 µm	± 0.025 µm	±	5.8E+00 µm
5 mm	4.94 mm	-60 µm	± 0.025 µm	±	5.8E+00 µm
15 mm	14.95 mm	-50 µm	± 0.074 µm	±	5.8E+00 µm
20 mm	20.00 mm	0 µm	± 0.100 µm	±	5.8E+00 µm
25 mm	25.02 mm	20 µm	± 0.125 µm	±	5.8E+00 µm