

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 0 °C              | 0.040 °C          | 0.0 °C    | -0.040 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 20 °C             | 20.035 °C         | 19.9 °C   | -0.135 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 35 °C             | 35.076 °C         | 35.1 °C   | 0.024 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.059 %HR        | 31.9 %HR  | 1.841 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 55 %HR            | 55.570 %HR        | 54.9 %HR  | -0.670 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 80 %HR            | 80.366 %HR        | 78.9 %HR  | -1.466 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |