

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.040 °C	0.1 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.9 °C	-0.135 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	34.9 °C	-0.176 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	31.4 %HR	1.341 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.570 %HR	54.6 %HR	-0.970 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.366 %HR	77.7 %HR	-2.666 %HR	± 3.0 %HR	± 1.9E+00 %HR