

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.040 °C	0.0 °C	-0.040 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.0 °C	-0.035 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.0 °C	-0.076 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	31.2 %HR	1.141 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.570 %HR	55.0 %HR	-0.570 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.366 %HR	80.3 %HR	-0.066 %HR	± 3.0 %HR	± 1.9E+00 %HR