

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.040 °C	-0.1 °C	-0.140 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.7 °C	-0.335 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	34.7 °C	-0.376 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.059 %HR	32.1 %HR	2.041 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.570 %HR	57.4 %HR	1.830 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.366 %HR	78.4 %HR	-1.966 %HR	± 3.0 %HR	± 1.9E+00 %HR