

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.015 °C	2.1 °C	0.085 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.8 °C	-0.193 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.152 °C	8.2 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C