

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.015 °C	2.0 °C	-0.015 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.9 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.152 °C	8.0 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C