

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-30 °C	-30.504 °C	-30.7 °C	-0.196 °C	± 0.5 °C	± 4.0E-01 °C
-20 °C	-20.597 °C	-20.5 °C	0.097 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.319 °C	-10.2 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	0.040 °C	0.2 °C	0.160 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.3 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.0 °C	-0.076 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.284 %HR	41.00 %HR	0.716 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.212 %HR	54.72 %HR	-0.492 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	77.58 %HR	2.327 %HR	± 3.0 %HR	± 1.9E+00 %HR