

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
59.931 s	60 s	0.069 s	0.1151%	± 0.0556%	±	2.9E-01 s
59.889 s	60 s	0.111 s	0.1853%	± 0.0556%	±	2.9E-01 s
59.965 s	60 s	0.035 s	0.0584%	± 0.0556%	±	2.9E-01 s