

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.112 °C	29.0 °C	-1.112 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	34.0 %HR	3.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.386 %HR	48.0 %HR	-2.386 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	69.0 %HR	-1.039 %HR	± 5.0 %HR	± 1.9E+00 %HR