

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.112 °C                  | 29.7 °C            | -0.412 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.570 %HR                 | 27.8 %HR           | -2.770 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.386 %HR                 | 52.7 %HR           | 2.314 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.039 %HR                 | 69.6 %HR           | -0.439 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |