

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
18.9281 l	18.9271 l	-0.0010 l	± 0.0038 l	±	7.7E-01 µl
5.0003 gal	5.0000 gal	-0.0003 gal	± 0.0010 gal	±	2.0E-07 gal