

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
100 mm	100.00 mm	-4 µm	± 5 µm	±	5.8E+00 µm
103 mm	103.00 mm	-4 µm	± 6 µm	±	5.8E+00 µm
108 mm	108.00 mm	-4 µm	± 6 µm	±	5.8E+00 µm
110 mm	110.00 mm	1 µm	± 6 µm	±	5.8E+00 µm
113 mm	113.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
115 mm	115.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
118 mm	118.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
120 mm	120.00 mm	2 µm	± 6 µm	±	5.8E+00 µm
123 mm	123.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
125 mm	125.00 mm	0 µm	± 6 µm	±	5.8E+00 µm