

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25 mm	25.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
28 mm	28.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
30 mm	30.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
35 mm	35.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
38 mm	38.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
40 mm	40.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
43 mm	43.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
45 mm	45.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
48 mm	48.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
50 mm	50.00 mm	0 μm	± 4 μm	±	5.8E+00 μm