

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 50 mm             | 50.00 mm  | 0 µm          | ± 4 µm            | ±                           | 5.8E+00 µm |
| 53 mm             | 53.00 mm  | -3 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 58 mm             | 58.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 60 mm             | 60.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 63 mm             | 63.00 mm  | -3 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 65 mm             | 65.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 68 mm             | 68.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 70 mm             | 70.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 73 mm             | 73.00 mm  | -2 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |
| 75 mm             | 75.00 mm  | -3 µm         | ± 5 µm            | ±                           | 5.8E+00 µm |