

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
75 mm	75.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
78 mm	78.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
80 mm	79.99 mm	-10 µm	± 5 µm	±	5.8E+00 µm
83 mm	83.00 mm	-3 µm	± 5 µm	±	5.8E+00 µm
85 mm	85.00 mm	-2 µm	± 5 µm	±	5.8E+00 µm
88 mm	88.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
90 mm	90.00 mm	-1 µm	± 5 µm	±	5.8E+00 µm
93 mm	93.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
95 mm	94.99 mm	-10 µm	± 5 µm	±	5.8E+00 µm
100 mm	100.00 mm	-1 µm	± 5 µm	±	5.8E+00 µm