

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0 mm	0.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
2 mm	2.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
5 mm	5.02 mm	0 µm	± 4 µm	±	5.8E+00 µm
10 mm	10.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
12 mm	12.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
15 mm	15.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
18 mm	18.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
20 mm	20.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
22 mm	22.02 mm	20 µm	± 4 µm	±	5.8E+00 µm
25 mm	25.02 mm	20 µm	± 4 µm	±	5.8E+00 µm