

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0 mm	0.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
2 mm	2.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
5 mm	5.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
10 mm	10.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
12 mm	12.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
15 mm	15.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
18 mm	18.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
20 mm	20.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
22 mm	22.02 mm	20 μm	± 4 μm	±	5.8E+00 μm
25 mm	25.02 mm	20 μm	± 4 μm	±	5.8E+00 μm