

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	# 0 mm	0.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	20 mm	20.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	40 mm	40.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	60 mm	60.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	80 mm	80.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	100 mm	100.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	120 mm	120.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	140 mm	140.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	160 mm	160.00 mm	0 µm	± 30 µm	± 1.2E+01 µm
Profundidad / (Depth)	200 mm	200.00 mm	0 µm	± 30 µm	± 1.2E+01 µm

