

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.9746 ml	2 ml	0.0254 ml	± 0.2 ml	±	1.5E+01 µl
4.9757 ml	5 ml	0.0243 ml	± 0.2 ml	±	6.2E+00 µl
9.9706 ml	10 ml	0.0294 ml	± 0.2 ml	±	5.8E+00 µl