

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 1.9775 ml                | 2 ml             | 0.0225 ml            | ± 0.2 ml                    | ±                                  | 6.4E+00 µl |
| 4.9744 ml                | 5 ml             | 0.0256 ml            | ± 0.2 ml                    | ±                                  | 5.2E+01 µl |
| 9.9667 ml                | 10 ml            | 0.0333 ml            | ± 0.2 ml                    | ±                                  | 7.3E+00 µl |