

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.9882 ml	2 ml	0.0118 ml	± 0.2 ml	±	6.3E+00 µl
4.9683 ml	5 ml	0.0317 ml	± 0.2 ml	±	2.8E+00 µl
9.9425 ml	10 ml	0.0575 ml	± 0.2 ml	±	1.0E+01 µl