

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.9880 ml	2 ml	0.0120 ml	± 0.2 ml	±	8.1E+00 µl
4.9657 ml	5 ml	0.0343 ml	± 0.2 ml	±	1.8E+01 µl
9.9456 ml	10 ml	0.0544 ml	± 0.2 ml	±	5.3E+00 µl