

Certificado No. (Certificate #):62700Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-10-22Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.0500 in	0.0500 in	0.0000 in	± 0.0001 in	±	3.8E-05 in
0.1000 in	0.1000 in	0.0000 in	± 0.0001 in	±	3.6E-05 in
0.2000 in	0.2000 in	0.0000 in	± 0.0001 in	±	3.5E-05 in
0.3000 in	0.3000 in	0.0000 in	± 0.0001 in	±	3.4E-05 in
0.4000 in	0.4000 in	0.0000 in	± 0.0001 in	±	4.1E-05 in
0.5000 in	0.5000 in	0.0000 in	± 0.0001 in	±	1.1E-04 in
0.6000 in	0.6000 in	0.0000 in	± 0.0001 in	±	4.3E-05 in
0.7000 in	0.7000 in	0.0000 in	± 0.0001 in	±	3.8E-05 in
0.8000 in	0.8000 in	0.0000 in	± 0.0001 in	±	4.2E-05 in
0.9000 in	0.9000 in	0.0000 in	± 0.0001 in	±	4.2E-05 in
1.0000 in	1.0000 in	0.0000 in	± 0.0001 in	±	3.5E-05 in