

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.062 °C         | 29.6 °C   | -0.462 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.570 %HR        | 26.1 %HR  | -4.470 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.335 %HR        | 47.0 %HR  | -3.335 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.345 %HR        | 65.8 %HR  | -4.545 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |