

Certificado No. (Certificate #):62713Equipo (Instrument):INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):2025-10-23Marca (Brand):Utest

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.000 mm	1.004 mm	4 µm		±	5.8E-01 µm
2.000 mm	2.004 mm	4 µm		±	5.8E-01 µm
4.000 mm	4.004 mm	4 µm		±	5.9E-01 µm
6.000 mm	6.003 mm	3 µm		±	6.1E-01 µm
8.000 mm	8.005 mm	5 µm		±	6.4E-01 µm
10.000 mm	10.012 mm	12 µm		±	6.7E-01 µm
12.000 mm	12.008 mm	8 µm		±	7.1E-01 µm
14.000 mm	14.030 mm	30 µm		±	7.5E-01 µm
15.000 mm	15.013 mm	13 µm		±	7.7E-01 µm