

Certificado No. (Certificate #):		62715	Equipo (Instrument):		INDICADOR DE DIAL
Fecha de Calibración (Calibration Date):		2025-10-23	Marca (Brand):		Novotechnik
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
1.000 mm		1.033 mm	33 µm	±	5.8E-01 µm
2.000 mm		2.021 mm	21 µm	±	5.8E-01 µm
4.000 mm		4.022 mm	22 µm	±	5.9E-01 µm
6.000 mm		6.008 mm	8 µm	±	6.1E-01 µm
8.000 mm		8.006 mm	6 µm	±	6.4E-01 µm
10.000 mm		10.015 mm	15 µm	±	6.7E-01 µm
15.000 mm		14.999 mm	-1 µm	±	7.7E-01 µm
20.000 mm		19.993 mm	-8 µm	±	8.9E-01 µm
25.000 mm		25.031 mm	31 µm	±	1.0E+00 µm