

Certificado No. (Certificate #):

62716

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-10-23

Marca (Brand):

Utest

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.000 mm	1.000 mm	0 µm		±	5.8E-01 µm
1.500 mm	1.497 mm	-3 µm		±	5.8E-01 µm
1.800 mm	1.801 mm	1 µm		±	5.8E-01 µm
2.000 mm	1.998 mm	-2 µm		±	5.8E-01 µm
5.000 mm	4.992 mm	-8 µm		±	6.0E-01 µm
6.000 mm	5.998 mm	-2 µm		±	6.1E-01 µm
7.000 mm	6.992 mm	-8 µm		±	6.2E-01 µm
8.000 mm	7.997 mm	-3 µm		±	6.4E-01 µm
10.000 mm	10.004 mm	4 µm		±	6.7E-01 µm