

Certificado No. (Certificate #):		62717	Equipo (Instrument): INDICADOR DE DIAL		
Fecha de Calibración (Calibration Date):		2025-10-23	Marca (Brand): Novotechnik		
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
1.000 mm		1.032 mm	32 µm	±	5.8E-01 µm
2.000 mm		2.006 mm	6 µm	±	5.8E-01 µm
4.000 mm		4.010 mm	10 µm	±	5.9E-01 µm
6.000 mm		6.021 mm	21 µm	±	6.1E-01 µm
8.000 mm		8.041 mm	40 µm	±	6.4E-01 µm
10.000 mm		10.052 mm	52 µm	±	6.7E-01 µm
15.000 mm		15.078 mm	78 µm	±	7.7E-01 µm
20.000 mm		20.093 mm	93 µm	±	8.9E-01 µm
25.000 mm		25.103 mm	103 µm	±	1.0E+00 µm