

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.855 °C	1.7 °C	-0.155 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.943 °C	4.8 °C	-0.143 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	7.781 °C	7.9 °C	0.119 °C	± 1.0 °C	± 4.0E-01 °C