

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.855 °C	2.2 °C	0.345 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.943 °C	5.2 °C	0.257 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	7.781 °C	8.0 °C	0.219 °C	± 1.0 °C	± 4.0E-01 °C