

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.481 °C	29.6 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.079 %HR	29.0 %HR	-1.079 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.828 %HR	49.9 %HR	-1.928 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.139 %HR	73.4 %HR	-0.739 %HR	± 3.0 %HR	± 1.9E+00 %HR