

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.450 °C	-25.4 °C	0.050 °C	± 1.0 °C	± 4.0E-01 °C
0 °C	-0.422 °C	-0.3 °C	0.122 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.394 °C	5.3 °C	-0.094 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.158 °C	10.2 °C	0.042 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.2 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.012 °C	29.9 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.090 °C	39.9 °C	-0.190 °C	± 0.5 °C	± 4.0E-01 °C

