

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.263 %HR	27.8 %HR	-2.463 %HR	± 4 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	48.8 %HR	-1.259 %HR	± 4 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	68.7 %HR	-1.236 %HR	± 4 %HR	± 1.9E+00 %HR
90 %HR	90.141 %HR	89.4 %HR	-0.741 %HR	± 4 %HR	± 1.9E+00 %HR

