

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.263 %HR	31.1 %HR	0.837 %HR	± 4 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	48.9 %HR	-1.159 %HR	± 4 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	68.9 %HR	-1.036 %HR	± 4 %HR	± 1.9E+00 %HR
90 %HR	90.141 %HR	92.6 %HR	2.459 %HR	± 4 %HR	± 1.9E+00 %HR

