

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.263 %HR	30.1 %HR	-0.163 %HR	± 4 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	46.7 %HR	-3.359 %HR	± 4 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	67.4 %HR	-2.536 %HR	± 4 %HR	± 1.9E+00 %HR
90 %HR	90.141 %HR	89.5 %HR	-0.641 %HR	± 4 %HR	± 1.9E+00 %HR

