

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.450 °C        | -25.7 °C  | -0.250 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.422 °C         | -0.5 °C   | -0.078 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.394 °C          | 5.3 °C    | -0.094 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.158 °C         | 10.1 °C   | -0.058 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.249 °C         | 25.7 °C   | 0.451 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.012 °C         | 30.4 °C   | 0.388 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.090 °C         | 40.3 °C   | 0.210 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

