

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.263 %HR	26.3 %HR	-3.963 %HR	± 4 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	46.8 %HR	-3.259 %HR	± 4 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	72.2 %HR	2.264 %HR	± 4 %HR	± 1.9E+00 %HR
90 %HR	90.141 %HR	89.3 %HR	-0.841 %HR	± 4 %HR	± 1.9E+00 %HR

