

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.450 °C        | -25.7 °C  | -0.250 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.422 °C         | -0.8 °C   | -0.378 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.394 °C          | 5.7 °C    | 0.306 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.158 °C         | 10.6 °C   | 0.442 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.249 °C         | 24.9 °C   | -0.349 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.012 °C         | 29.6 °C   | -0.412 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.090 °C         | 40.5 °C   | 0.410 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

