

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.263 %HR	27.7 %HR	-2.563 %HR	± 4 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	49.7 %HR	-0.359 %HR	± 4 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	69.8 %HR	-0.136 %HR	± 4 %HR	± 1.9E+00 %HR
90 %HR	90.141 %HR	88.6 %HR	-1.541 %HR	± 4 %HR	± 1.9E+00 %HR

