

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.153 °C	29.6 °C	-0.553 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	31.807 %HR	28.2 %HR	-3.607 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.908 %HR	49.2 %HR	-1.708 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.867 %HR	74.4 %HR	3.533 %HR	± 5.0 %HR	± 1.9E+00 %HR