

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	26.1 %HR	-4.470 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.335 %HR	47.8 %HR	-2.535 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	65.8 %HR	-4.545 %HR	± 3.0 %HR	± 1.9E+00 %HR