

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.282 °C	16.1 °C	0.818 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.179 °C	26.1 °C	0.921 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	50.7 °C	0.482 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.130 %HR	25.6 %HR	0.470 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	51.0 %HR	0.675 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	71.0 %HR	-4.253 %HR	± 5.0 %HR	± 1.9E+00 %HR