

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.282 °C	16.0 °C	0.718 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.179 °C	25.9 °C	0.721 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	50.2 °C	-0.018 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.130 %HR	33.0 %HR	7.870 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	57.0 %HR	6.675 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	82.0 %HR	6.747 %HR	± 5.0 %HR	± 1.9E+00 %HR