

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.282 °C	16.2 °C	0.918 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.179 °C	26.1 °C	0.921 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.218 °C	51.0 °C	0.782 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.130 %HR	28 %HR	2.870 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	53 %HR	2.675 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	77 %HR	1.747 %HR	± 5.0 %HR	± 1.9E+00 %HR