

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	50 mm	50.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	100 mm	99.99 mm	-10 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	130 mm	130.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	180 mm	180.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	230 mm	230.02 mm	20 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	250 mm	250.02 mm	20 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	300 mm	300.01 mm	10 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	350 mm	349.99 mm	-10 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	400 mm	400.01 mm	10 µm	± 40 µm	± 6.1E+00 µm